

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027300.D
 Acq On : 04 Mar 2022 14:44
 Operator : JC/MD
 Sample : N1760-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-1-030222

Quant Time: Mar 05 00:44:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	80275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	131386	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121634	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	57449	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47905	49.688	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.380%
35) Dibromofluoromethane	5.391	113	41282	50.620	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.240%
50) Toluene-d8	8.653	98	151881	49.807	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.620%
62) 4-Bromofluorobenzene	11.079	95	57092	48.775	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.560%
Target Compounds						
16) Acetone	2.380	43	989	2.336	ug/l	# 68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027300.D
Acq On : 04 Mar 2022 14:44
Operator : JC/MD
Sample : N1760-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-1-030222

Quant Time: Mar 05 00:44:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

