

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027302.D
 Acq On : 04 Mar 2022 15:30
 Operator : JC/MD
 Sample : N1793-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-3

Quant Time: Mar 05 00:44:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	76965	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	123337	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113573	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	45833	49.584	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	99.160%		
35) Dibromofluoromethane	5.391	113	39339	51.385	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	102.780%		
50) Toluene-d8	8.653	98	141411	49.400	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.800%		
62) 4-Bromofluorobenzene	11.079	95	52452	47.736	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.480%		
Target Compounds						
16) Acetone	2.380	43	1463	3.605	ug/l	95
52) Toluene	8.720	92	5946	2.842	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027302.D
Acq On : 04 Mar 2022 15:30
Operator : JC/MD
Sample : N1793-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-3

Quant Time: Mar 05 00:44:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

