

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027303.D
 Acq On : 04 Mar 2022 15:54
 Operator : JC/MD
 Sample : N1793-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-4

Quant Time: Mar 05 00:44:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	78016	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	126987	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	46566	49.698	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.400%
35) Dibromofluoromethane	5.391	113	39666	50.323	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.640%
50) Toluene-d8	8.653	98	143955	48.843	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	54544	48.213	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.420%
Target Compounds						
52) Toluene	8.720	92	3559	1.652	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027303.D
Acq On : 04 Mar 2022 15:54
Operator : JC/MD
Sample : N1793-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-4

Quant Time: Mar 05 00:44:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

