

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027304.D
 Acq On : 04 Mar 2022 16:17
 Operator : JC/MD
 Sample : N1793-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-5

Quant Time: Mar 05 00:45:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	78275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	125968	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	112142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	46630	49.602	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.200%
35) Dibromofluoromethane	5.391	113	40422	51.697	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.400%
50) Toluene-d8	8.647	98	150525	51.486	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.980%
62) 4-Bromofluorobenzene	11.079	95	52480	46.764	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.520%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12402	30.047	ug/l	98
20) Methylene Chloride	2.788	84	756	0.884	ug/l	88
25) 2-Butanone	4.568	43	5790	8.808	ug/l	93
29) Tetrahydrofuran	5.025	42	1567	3.562	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027304.D
Acq On : 04 Mar 2022 16:17
Operator : JC/MD
Sample : N1793-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-5

Quant Time: Mar 05 00:45:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

