

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027305.D
 Acq On : 04 Mar 2022 16:41
 Operator : JC/MD
 Sample : N1793-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-6

Quant Time: Mar 05 00:45:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	74978	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	124947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	113528	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53054	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	45566	50.601	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.200%
35) Dibromofluoromethane	5.391	113	39053	50.354	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.700%
50) Toluene-d8	8.647	98	142702	49.209	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.420%
62) 4-Bromofluorobenzene	11.079	95	53594	48.146	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.300%
Target Compounds						Qvalue
16) Acetone	2.380	43	808	2.044	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027305.D
Acq On : 04 Mar 2022 16:41
Operator : JC/MD
Sample : N1793-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-6

Quant Time: Mar 05 00:45:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

