

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027307.D
 Acq On : 04 Mar 2022 17:27
 Operator : JC/MD
 Sample : N1783-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Mar 05 00:45:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75192	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	123501	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	110492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51490	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	45528	50.415	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 100.840%	
35) Dibromofluoromethane	5.391	113	39216	51.156	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.320%	
50) Toluene-d8	8.647	98	144183	50.301	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.600%	
62) 4-Bromofluorobenzene	11.079	95	52284	47.520	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.040%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	847	2.136	ug/l	100
27) cis-1,2-Dichloroethene	4.489	96	2302	2.543	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027307.D
Acq On : 04 Mar 2022 17:27
Operator : JC/MD
Sample : N1783-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Mar 05 00:45:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

