

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027312.D
 Acq On : 04 Mar 2022 19:25
 Operator : JC/MD
 Sample : N1783-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Mar 05 00:46:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	69753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	115226	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105972	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	42785	51.072	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 102.140%	
35) Dibromofluoromethane	5.391	113	36745	51.375	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.760%	
50) Toluene-d8	8.653	98	131589	49.205	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.400%	
62) 4-Bromofluorobenzene	11.085	95	50205	48.907	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.820%	
Target Compounds						
16) Acetone	2.380	43	888	2.414	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027312.D
Acq On : 04 Mar 2022 19:25
Operator : JC/MD
Sample : N1783-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Mar 05 00:46:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

