

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027314.D
 Acq On : 04 Mar 2022 20:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 05 00:46:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	70	0.00
2 T	Dichlorodifluoromethane	50.000	44.616	10.8	64	0.00
3 P	Chloromethane	50.000	39.355	21.3	55	0.00
4 C	Vinyl Chloride	50.000	39.473	21.1#	54	0.00
5 T	Bromomethane	50.000	39.782	20.4	61	0.00
6 T	Chloroethane	50.000	41.155	17.7	59	0.00
7 T	Trichlorofluoromethane	50.000	49.346	1.3	69	0.00
8 T	Diethyl Ether	50.000	43.324	13.4	59	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	47.824	4.4	67	0.00
10 T	Methyl Iodide	50.000	41.415	17.2	55	0.00
11 T	Tert butyl alcohol	250.000	227.967	8.8	65	0.00
12 CM	1,1-Dichloroethene	50.000	45.999	8.0#	64	0.00
13 T	Acrolein	250.000	234.972	6.0	68	0.00
14 T	Allyl chloride	50.000	43.391	13.2	60	0.00
15 T	Acrylonitrile	250.000	241.116	3.6	65	0.00
16 T	Acetone	250.000	229.815	8.1	66	0.00
17 T	Carbon Disulfide	50.000	41.201	17.6	58	0.00
18 T	Methyl Acetate	50.000	45.278	9.4	62	0.00
19 T	Methyl tert-butyl Ether	50.000	49.064	1.9	67	0.00
20 T	Methylene Chloride	50.000	47.600	4.8	67	0.00
21 T	trans-1,2-Dichloroethene	50.000	46.938	6.1	65	0.00
22 T	Diisopropyl ether	50.000	45.366	9.3	62	0.00
23 T	Vinyl Acetate	250.000	238.643	4.5	63	0.00
24 P	1,1-Dichloroethane	50.000	47.422	5.2	66	0.00
25 T	2-Butanone	250.000	232.722	6.9	64	0.00
26 T	2,2-Dichloropropane	50.000	49.032	1.9	67	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.834	0.3	68	0.00
28 T	Bromochloromethane	50.000	50.148	-0.3	68	0.00
29 T	Tetrahydrofuran	250.000	224.266	10.3	62	0.00
30 C	Chloroform	50.000	49.883	0.2#	69	0.00
31 T	Cyclohexane	50.000	43.203	13.6	60	0.00
32 T	1,1,1-Trichloroethane	50.000	51.774	-3.5	70	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.153	-6.3	77	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	69	0.00
35 S	Dibromofluoromethane	50.000	53.528	-7.1	75	0.00
36 T	1,1-Dichloropropene	50.000	48.099	3.8	65	0.00
37 T	Ethyl Acetate	50.000	46.842	6.3	64	0.00
38 T	Carbon Tetrachloride	50.000	52.946	-5.9	70	0.00
39 T	Methylcyclohexane	50.000	47.385	5.2	64	0.00
40 TM	Benzene	50.000	48.844	2.3	66	0.00
41 T	Methacrylonitrile	50.000	48.313	3.4	65	0.00
42 TM	1,2-Dichloroethane	50.000	52.315	-4.6	70	0.00
43 T	Isopropyl Acetate	50.000	48.554	2.9	64	0.00
44 TM	Trichloroethene	50.000	50.321	-0.6	68	0.00
45 C	1,2-Dichloropropane	50.000	49.574	0.9#	65	0.00
46 T	Dibromomethane	50.000	52.113	-4.2	69	0.00
47 T	Bromodichloromethane	50.000	53.232	-6.5	72	0.00
48 T	Methyl methacrylate	50.000	50.279	-0.6	65	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
 Data File : VX027314.D
 Acq On : 04 Mar 2022 20:11
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 05 00:46:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 23 13:06:57 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	992.126	0.8	66	0.00
50 S	Toluene-d8	50.000	51.665	-3.3	73	0.00
51 T	4-Methyl-2-Pentanone	250.000	251.080	-0.4	66	0.00
52 CM	Toluene	50.000	51.086	-2.2#	69	0.00
53 T	t-1,3-Dichloropropene	50.000	47.141	5.7	68	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.492	-9.0	70	0.00
55 T	1,1,2-Trichloroethane	50.000	53.244	-6.5	71	0.00
56 T	Ethyl methacrylate	50.000	51.426	-2.9	66	0.00
57 T	1,3-Dichloropropane	50.000	51.768	-3.5	68	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	247.458	1.0	65	0.00
59 T	2-Hexanone	250.000	247.106	1.2	65	0.00
60 T	Dibromochloromethane	50.000	57.273	-14.5	74	0.00
61 T	1,2-Dibromoethane	50.000	52.695	-5.4	69	0.00
62 S	4-Bromofluorobenzene	50.000	53.188	-6.4	74	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	70	0.00
64 T	Tetrachloroethene	50.000	46.335	7.3	63	0.00
65 PM	Chlorobenzene	50.000	50.588	-1.2	69	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.402	-8.8	74	0.00
67 C	Ethyl Benzene	50.000	50.430	-0.9#	68	0.00
68 T	m/p-Xylenes	100.000	101.961	-2.0	68	0.00
69 T	o-Xylene	50.000	50.208	-0.4	68	0.00
70 T	Styrene	50.000	51.852	-3.7	68	0.00
71 P	Bromoform	50.000	50.560	-1.1	75	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	69	0.00
73 T	Isopropylbenzene	50.000	50.912	-1.8	69	0.00
74 T	N-amyl acetate	50.000	51.839	-3.7	67	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	52.464	-4.9	73	0.00
76 T	1,2,3-Trichloropropane	50.000	51.953	-3.9	70	0.00
77 T	Bromobenzene	50.000	51.663	-3.3	70	0.00
78 T	n-propylbenzene	50.000	50.375	-0.8	68	0.00
79 T	2-Chlorotoluene	50.000	50.553	-1.1	69	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.799	-5.6	71	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	45.162	9.7	66	0.00
82 T	4-Chlorotoluene	50.000	50.843	-1.7	69	0.00
83 T	tert-Butylbenzene	50.000	54.019	-8.0	73	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.382	-4.8	70	0.00
85 T	sec-Butylbenzene	50.000	51.698	-3.4	69	0.00
86 T	p-Isopropyltoluene	50.000	52.510	-5.0	70	0.00
87 T	1,3-Dichlorobenzene	50.000	51.335	-2.7	70	0.00
88 T	1,4-Dichlorobenzene	50.000	50.078	-0.2	69	0.00
89 T	n-Butylbenzene	50.000	50.826	-1.7	67	0.00
90 T	Hexachloroethane	50.000	54.630	-9.3	71	0.00
91 T	1,2-Dichlorobenzene	50.000	51.529	-3.1	70	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.676	-9.4	72	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.745	-5.5	70	0.00
94 T	Hexachlorobutadiene	50.000	52.442	-4.9	71	0.00
95 T	Naphthalene	50.000	52.790	-5.6	68	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030422\
Data File : VX027314.D
Acq On : 04 Mar 2022 20:11
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 05 00:46:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022322W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 23 13:06:57 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	53.538	-7.1	71	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6