

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030424\
 Data File : VX040565.D
 Acq On : 04 Mar 2024 13:47
 Operator : JC/MD
 Sample : PB159402TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB159402TB

Quant Time: Mar 05 00:24:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	75584	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	128080	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	60634	47.963	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.920%
35) Dibromofluoromethane	5.385	113	42616	47.848	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.700%
50) Toluene-d8	8.647	98	160279	49.576	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.160%
62) 4-Bromofluorobenzene	11.079	95	58213	47.236	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	94.480%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	1131	5.322	ug/l #	92
16) Acetone	2.380	43	33411	81.041	ug/l	97
20) Methylene Chloride	2.788	84	17426	17.663	ug/l	98
43) Isopropyl Acetate	6.342	43	69573	29.524	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030424\
Data File : VX040565.D
Acq On : 04 Mar 2024 13:47
Operator : JC/MD
Sample : PB159402TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB159402TB

Quant Time: Mar 05 00:24:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

