

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030424\
 Data File : VX040567.D
 Acq On : 04 Mar 2024 14:34
 Operator : JC/MD
 Sample : P1662-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RBR-2520556

Quant Time: Mar 05 00:26:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 27 05:01:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	73613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	118559	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128656	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59941	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	61118	49.640	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.280%
35) Dibromofluoromethane	5.391	113	43155	52.344	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.680%
50) Toluene-d8	8.647	98	146130	48.830	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.660%
62) 4-Bromofluorobenzene	11.079	95	61326	53.758	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	107.520%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	50010	124.551	ug/l	99
20) Methylene Chloride	2.794	84	12087	12.579	ug/l	91
25) 2-Butanone	4.574	43	3778	5.400	ug/l	97
37) Ethyl Acetate	4.727	43	1876	1.350	ug/l #	85
43) Isopropyl Acetate	6.336	43	54507	24.989	ug/l	98
95) Naphthalene	13.780	128	5295	1.281	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030424\
Data File : VX040567.D
Acq On : 04 Mar 2024 14:34
Operator : JC/MD
Sample : P1662-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RBR-2520556

Quant Time: Mar 05 00:26:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022624W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 27 05:01:33 2024
Response via : Initial Calibration

