

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030425\  
 Data File : VX045122.D  
 Acq On : 04 Mar 2025 10:04  
 Operator : JC/MD  
 Sample : VX0304WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 5 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0304WBL01

Quant Time: Mar 05 00:46:24 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X022125W.M  
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
 QLast Update : Sat Feb 22 01:06:36 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Bromochloromethane       | 4.897  | 128   | 14872    | 30.000   | ug/l  | 0.01     |
| 28) 1,4-Difluorobenzene     | 6.757  | 114   | 78579    | 30.000   | ug/l  | 0.00     |
| 57) Chlorobenzene-d5        | 10.049 | 117   | 69103    | 30.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 27) 1,2-Dichloroethane-d4   | 5.946  | 65    | 43219    | 29.921   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 110 | Recovery | =     | 99.733%  |
| 60) 4-Bromofluorobenzene    | 11.079 | 95    | 35500    | 27.485   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 63 - 112 | Recovery | =     | 91.600%  |
| 63) Toluene-d8              | 8.647  | 98    | 115602   | 30.239   | ug/l  | 0.00     |
| Spiked Amount               | 30.000 | Range | 91 - 112 | Recovery | =     | 100.800% |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030425\  
Data File : VX045122.D  
Acq On : 04 Mar 2025 10:04  
Operator : JC/MD  
Sample : VX0304WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 5 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0304WBL01

Quant Time: Mar 05 00:46:24 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\624X022125W.M  
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS  
QLast Update : Sat Feb 22 01:06:36 2025  
Response via : Initial Calibration

