

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030519\
 Data File : VX007835.D
 Acq On : 05 Mar 2019 14:21
 Operator : JC/SP
 Sample : K1659-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 M1-T5-OZ(PRE)

Quant Time: Mar 06 12:33:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	313879	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	487671	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	459450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	219012	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	176440	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
35) Dibromofluoromethane	5.50	113	161436	49.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.04%	
50) Toluene-d8	8.71	98	590067	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
62) 4-Bromofluorobenzene	11.13	95	213145	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
Target Compounds						
16) Acetone	2.45	43	5363	3.102	ug/l	98
24) 1,1-Dichloroethane	3.70	63	5612	0.977	ug/l	99
29) Tetrahydrofuran	5.16	42	2052	1.182	ug/l #	74
44) Trichloroethene	7.21	130	2870	0.771	ug/l	97
64) Tetrachloroethene	9.33	164	1818	0.487	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030519\
Data File : VX007835.D
Acq On : 05 Mar 2019 14:21
Operator : JC/SP
Sample : K1659-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
M1-T5-OZ(PRE)

Quant Time: Mar 06 12:33:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

