

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030519\
 Data File : VX007847.D
 Acq On : 05 Mar 2019 19:41
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 06 01:10:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 01:53:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	296634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	462082	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	430717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	207700	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	164234	44.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.40%	
35) Dibromofluoromethane	5.51	113	155259	50.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.52%	
50) Toluene-d8	8.71	98	555982	47.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.04%	
62) 4-Bromofluorobenzene	11.13	95	202078	46.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030519\
Data File : VX007847.D
Acq On : 05 Mar 2019 19:41
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 06 01:10:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 01:53:13 2019
Response via : Initial Calibration

