

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030520\
 Data File : VX015097.D
 Acq On : 05 Mar 2020 09:19
 Operator : JC/SP
 Sample : VX0305WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0305WBL01

Quant Time: Mar 06 03:32:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	364113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	592052	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	547977	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	249877	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	220916	51.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.28%	
35) Dibromofluoromethane	5.48	113	182973	48.99	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.98%	
50) Toluene-d8	8.71	98	712224	50.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.26%	
62) 4-Bromofluorobenzene	11.13	95	240611	47.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.96%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030520\
Data File : VX015097.D
Acq On : 05 Mar 2020 09:19
Operator : JC/SP
Sample : VX0305WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0305WBL01

Quant Time: Mar 06 03:32:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

