

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030520\
 Data File : VX015107.D
 Acq On : 05 Mar 2020 13:26
 Operator : JC/SP
 Sample : L1812-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VOC-DRUM-030420

Quant Time: Mar 06 03:42:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	362017	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	575923	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	532461	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	270844	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	210033	48.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.80%	
35) Dibromofluoromethane	5.50	113	158729	43.69	ug/l	0.01
Spiked Amount 50.000			Recovery	=	87.38%	
50) Toluene-d8	8.71	98	696594	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
62) 4-Bromofluorobenzene	11.14	95	248979	50.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.02%	
Target Compounds						
3) Chloromethane	1.32	50	19103	4.254	ug/l	93
16) Acetone	2.50	43	379869	218.398	ug/l	97
18) Methyl Acetate	2.80	43	5216135	1226.460	ug/l	99
29) Tetrahydrofuran	5.17	42	53292	32.467	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030520\
Data File : VX015107.D
Acq On : 05 Mar 2020 13:26
Operator : JC/SP
Sample : L1812-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VOC-DRUM-030420

Quant Time: Mar 06 03:42:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

