

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030520\
 Data File : VX015114.D
 Acq On : 05 Mar 2020 17:27
 Operator : JC/SP
 Sample : L1818-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-4A

Quant Time: Mar 06 04:00:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 04 14:19:10 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	346681	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	571444	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	551355	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	290543	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	211130	51.33	ug/l	0.00
Spiked Amount			Recovery	=	102.66%	
35) Dibromofluoromethane	5.49	113	174653	48.45	ug/l	0.00
Spiked Amount			Recovery	=	96.90%	
50) Toluene-d8	8.71	98	687321	50.12	ug/l	0.00
Spiked Amount			Recovery	=	100.24%	
62) 4-Bromofluorobenzene	11.13	95	257587	52.66	ug/l	0.00
Spiked Amount			Recovery	=	105.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030520\
Data File : VX015114.D
Acq On : 05 Mar 2020 17:27
Operator : JC/SP
Sample : L1818-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-4A

Quant Time: Mar 06 04:00:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030420W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 04 14:19:10 2020
Response via : Initial Calibration

