

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030524\
 Data File : VX040584.D
 Acq On : 05 Mar 2024 12:53
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 06 00:09:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 05 13:21:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	14461	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	78791	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	69001	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	41767	28.919	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.400%
60) 4-Bromofluorobenzene	11.079	95	34794	29.397	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	98.000%
63) Toluene-d8	8.647	98	98399	32.269	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	107.567%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030524\
Data File : VX040584.D
Acq On : 05 Mar 2024 12:53
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 06 00:09:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 05 13:21:29 2024
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.897 min Scan# 61
Delta R.T. -0.001 min
Lab File: VX040584.D
Acq: 05 Mar 2024 12:53




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:128 Resp: 14461
Ion Ratio Lower Upper
128 100
49 170.5 0.0 444.0
130 128.9 0.0 297.8



#27
1,2-Dichloroethane-d4
Concen: 28.919 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.006 min
Lab File: VX040584.D
Acq: 05 Mar 2024 12:53

Tgt Ion: 65 Resp: 41767
Ion Ratio Lower Upper
65 100
67 50.6 42.5 63.7









#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. 0.006 min
Lab File: VX040584.D
Acq: 05 Mar 2024 12:53




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 78791
Ion Ratio Lower Upper
114 100
63 23.0 17.0 25.4
88 17.1 12.7 19.1



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX040584.D
Acq: 05 Mar 2024 12:53

Tgt Ion:117 Resp: 69001
Ion Ratio Lower Upper
117 100
82 59.8 47.0 70.6
119 32.3 25.2 37.8









#60
4-Bromofluorobenzene
Concen: 29.397 ug/l
RT: 11.079 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX040584.D
Acq: 05 Mar 2024 12:53

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion: 95 Resp: 34794
Ion Ratio Lower Upper
95 100
174 75.2 55.4 83.2
176 72.2 52.6 78.8



#63
Toluene-d8
Concen: 32.269 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX040584.D
Acq: 05 Mar 2024 12:53

Tgt Ion: 98 Resp: 98399
Ion Ratio Lower Upper
98 100
100 63.2 51.0 76.4

