

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045133.D
 Acq On : 05 Mar 2025 11:40
 Operator : JC/MD
 Sample : Q1448-04ME
 Misc : 6.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P3ME

Quant Time: Mar 06 01:02:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	67331	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128533	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115484	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55033	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	52674	49.182	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.360%
35) Dibromofluoromethane	5.379	113	42058	48.935	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.860%
50) Toluene-d8	8.647	98	157997	50.707	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.420%
62) 4-Bromofluorobenzene	11.079	95	56403	54.627	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	109.260%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	1522	3.063	ug/l	96
31) Cyclohexane	5.452	56	229490	157.496	ug/l	96
39) Methylcyclohexane	7.373	83	305736	216.477	ug/l	98
40) Benzene	6.031	78	70701	18.996	ug/l	98
52) Toluene	8.720	92	3128	1.432	ug/l	90
68) m/p-Xylenes	10.305	106	2532	1.624	ug/l	83
69) o-Xylene	10.640	106	3642	2.317	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045133.D
Acq On : 05 Mar 2025 11:40
Operator : JC/MD
Sample : Q1448-04ME
Misc : 6.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P3ME

Quant Time: Mar 06 01:02:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

