

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045143.D
 Acq On : 05 Mar 2025 15:32
 Operator : JC/MD
 Sample : Q1478-08 100X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-AQ-DRUM-612-022825

Quant Time: Mar 06 01:08:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	77819	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152987	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	143136	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65197	52.671	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.340%
35) Dibromofluoromethane	5.385	113	53117	51.924	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.840%
50) Toluene-d8	8.647	98	192696	51.958	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.920%
62) 4-Bromofluorobenzene	11.079	95	66421	54.047	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.100%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	1743	3.035	ug/l	92
27) cis-1,2-Dichloroethene	4.501	96	1341	1.150	ug/l	96
29) Tetrahydrofuran	5.019	42	1256	2.163	ug/l #	70
44) Trichloroethene	7.129	130	6958	6.689	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045143.D
Acq On : 05 Mar 2025 15:32
Operator : JC/MD
Sample : Q1478-08 100X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-AQ-DRUM-612-022825

Quant Time: Mar 06 01:08:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

