

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045146.D
 Acq On : 05 Mar 2025 16:42
 Operator : JC/MD
 Sample : Q1501-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB01-030525

Quant Time: Mar 06 01:09:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	76618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152804	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	61013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64818	53.185	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.380%	
35) Dibromofluoromethane	5.379	113	52313	51.199	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.400%	
50) Toluene-d8	8.647	98	192904	52.077	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.160%	
62) 4-Bromofluorobenzene	11.079	95	65419	53.296	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.600%	
Target Compounds						
11) Tert butyl alcohol	2.971	59	2345	10.158	ug/l #	87
16) Acetone	2.380	43	1127	1.993	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045146.D
Acq On : 05 Mar 2025 16:42
Operator : JC/MD
Sample : Q1501-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB01-030525

Quant Time: Mar 06 01:09:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

