

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045151.D
 Acq On : 05 Mar 2025 18:38
 Operator : JC/MD
 Sample : Q1488-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ENV-104-GW01

Quant Time: Mar 06 01:11:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	79416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	156296	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	141923	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59794	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65462	51.821	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.640%
35) Dibromofluoromethane	5.379	113	53404	51.099	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.200%
50) Toluene-d8	8.646	98	193262	51.008	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.020%
62) 4-Bromofluorobenzene	11.079	95	64393	51.288	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.580%
Target Compounds						Qvalue
16) Acetone	2.385	43	485114	827.700	ug/l	99
31) Cyclohexane	5.458	56	2244	1.306	ug/l	98
39) Methylcyclohexane	7.372	83	675	0.393	ug/l #	85
52) Toluene	8.720	92	822	0.310	ug/l #	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045151.D
Acq On : 05 Mar 2025 18:38
Operator : JC/MD
Sample : Q1488-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ENV-104-GW01

Quant Time: Mar 06 01:11:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

