

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045127.D
 Acq On : 05 Mar 2025 09:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 00:58:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.629	0.611	2.9	106	0.00
3 P	Chloromethane	0.770	0.671	12.9	97	0.00
4 C	Vinyl Chloride	0.764	0.689	9.8#	99	0.00
5 T	Bromomethane	0.300	0.271	9.7	101	0.00
6 T	Chloroethane	0.352	0.291	17.3	85	0.00
7 T	Trichlorofluoromethane	1.011	0.938	7.2	97	0.00
8 T	Diethyl Ether	0.396	0.348	12.1	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.581	0.581	0.0	106	0.00
10 T	Methyl Iodide	0.726	0.698	3.9	99	0.00
11 T	Tert butyl alcohol	0.151	0.108	28.5#	87	0.01
12 CM	1,1-Dichloroethene	0.614	0.553	9.9#	96	0.00
13 T	Acrolein	0.174	0.167	4.0	104	0.00
14 T	Allyl chloride	1.094	1.041	4.8	102	0.00
15 T	Acrylonitrile	0.404	0.349	13.6	93	0.00
16 T	Acetone	0.369	0.301	18.4	93	0.00
17 T	Carbon Disulfide	1.636	1.452	11.2	95	0.00
18 T	Methyl Acetate	0.888	0.810	8.8	102	0.00
19 T	Methyl tert-butyl Ether	2.061	1.855	10.0	97	0.00
20 T	Methylene Chloride	0.731	0.640	12.4	100	0.00
21 T	trans-1,2-Dichloroethene	0.607	0.571	5.9	98	0.00
22 T	Diisopropyl ether	2.224	2.073	6.8	98	0.00
23 T	Vinyl Acetate	1.856	1.760	5.2	97	0.00
24 P	1,1-Dichloroethane	1.247	1.123	9.9	98	0.00
25 T	2-Butanone	0.555	0.488	12.1	92	0.00
26 T	2,2-Dichloropropane	0.585	0.795	-35.9#	145	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.684	8.7	98	0.00
28 T	Bromochloromethane	0.594	0.559	5.9	101	0.00
29 T	Tetrahydrofuran	0.373	0.320	14.2	93	0.00
30 C	Chloroform	1.239	1.123	9.4#	98	0.00
31 T	Cyclohexane	1.082	1.035	4.3	102	0.00
32 T	1,1,1-Trichloroethane	1.000	0.923	7.7	98	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.743	6.5	107	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.334	0.344	-3.0	107	-0.01
36 T	1,1-Dichloropropene	0.459	0.452	1.5	99	-0.01
37 T	Ethyl Acetate	0.591	0.554	6.3	93	0.00
38 T	Carbon Tetrachloride	0.465	0.457	1.7	100	0.00
39 T	Methylcyclohexane	0.549	0.609	-10.9	106	0.00
40 TM	Benzene	1.448	1.416	2.2	97	0.00
41 T	Methacrylonitrile	0.319	0.314	1.6	94	0.00
42 TM	1,2-Dichloroethane	0.521	0.506	2.9	98	0.00
43 T	Isopropyl Acetate	0.874	0.847	3.1	93	0.00
44 TM	Trichloroethene	0.340	0.338	0.6	101	0.00
45 C	1,2-Dichloropropane	0.372	0.355	4.6#	98	0.00
46 T	Dibromomethane	0.263	0.255	3.0	97	0.00
47 T	Bromodichloromethane	0.516	0.512	0.8	100	0.00
48 T	Methyl methacrylate	0.438	0.423	3.4	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045127.D
 Acq On : 05 Mar 2025 09:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 00:58:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.008	11.1	90	0.00
50 S	Toluene-d8	1.212	1.229	-1.4	104	0.00
51 T	4-Methyl-2-Pentanone	0.587	0.550	6.3	92	0.00
52 CM	Toluene	0.849	0.858	-1.1#	98	0.00
53 T	t-1,3-Dichloropropene	0.431	0.464	-7.7	101	0.00
54 T	cis-1,3-Dichloropropene	0.503	0.541	-7.6	103	0.00
55 T	1,1,2-Trichloroethane	0.350	0.334	4.6	96	0.00
56 T	Ethyl methacrylate	0.532	0.519	2.4	93	0.00
57 T	1,3-Dichloropropane	0.605	0.579	4.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.276	-6.6	100	0.00
59 T	2-Hexanone	0.425	0.404	4.9	92	0.00
60 T	Dibromochloromethane	0.366	0.374	-2.2	100	0.00
61 T	1,2-Dibromoethane	0.349	0.340	2.6	98	0.00
62 S	4-Bromofluorobenzene	0.402	0.412	-2.5	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.320	0.328	-2.5	102	0.00
65 PM	Chlorobenzene	1.058	1.044	1.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.352	0.344	2.3	98	0.00
67 C	Ethyl Benzene	1.843	1.877	-1.8#	97	0.00
68 T	m/p-Xylenes	0.675	0.695	-3.0	97	0.00
69 T	o-Xylene	0.681	0.676	0.7	97	0.00
70 T	Styrene	1.101	1.118	-1.5	96	0.00
71 P	Bromoform	0.266	0.272	-2.3	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.903	3.893	0.3	98	0.00
74 T	N-amyl acetate	1.831	1.795	2.0	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.432	1.304	8.9	96	0.00
76 T	1,2,3-Trichloropropane	1.164	1.040	10.7	95	0.00
77 T	Bromobenzene	0.921	0.885	3.9	99	0.00
78 T	n-propylbenzene	4.409	4.545	-3.1	99	0.00
79 T	2-Chlorotoluene	2.808	2.644	5.8	96	0.00
80 T	1,3,5-Trimethylbenzene	3.156	3.166	-0.3	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.350	-2.9	108	0.00
82 T	4-Chlorotoluene	3.060	3.007	1.7	97	0.00
83 T	tert-Butylbenzene	3.242	3.126	3.6	97	0.00
84 T	1,2,4-Trimethylbenzene	3.175	3.172	0.1	98	0.00
85 T	sec-Butylbenzene	3.884	4.003	-3.1	99	0.00
86 T	p-Isopropyltoluene	3.139	3.249	-3.5	99	0.00
87 T	1,3-Dichlorobenzene	1.643	1.598	2.7	100	0.00
88 T	1,4-Dichlorobenzene	1.662	1.585	4.6	98	0.00
89 T	n-Butylbenzene	2.703	2.981	-10.3	103	0.00
90 T	Hexachloroethane	0.568	0.565	0.5	98	0.00
91 T	1,2-Dichlorobenzene	1.648	1.633	0.9	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.264	5.4	95	0.00
93 T	1,2,4-Trichlorobenzene	0.938	0.991	-5.7	103	0.00
94 T	Hexachlorobutadiene	0.385	0.405	-5.2	108	0.00
95 T	Naphthalene	3.605	3.622	-0.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045127.D
Acq On : 05 Mar 2025 09:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 06 00:58:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.990	1.016	-2.6	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6