

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045153.D
 Acq On : 05 Mar 2025 19:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 01:12:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.629	0.620	1.4	90	0.00
3 P	Chloromethane	0.770	0.726	5.7	88	0.00
4 C	Vinyl Chloride	0.764	0.737	3.5#	88	0.00
5 T	Bromomethane	0.300	0.281	6.3	88	0.00
6 T	Chloroethane	0.352	0.295	16.2	72	0.00
7 T	Trichlorofluoromethane	1.011	0.998	1.3	86	0.00
8 T	Diethyl Ether	0.396	0.381	3.8	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.581	0.603	-3.8	92	0.00
10 T	Methyl Iodide	0.726	0.774	-6.6	92	0.00
11 T	Tert butyl alcohol	0.151	0.122	19.2	83	0.01
12 CM	1,1-Dichloroethene	0.614	0.600	2.3#	87	0.00
13 T	Acrolein	0.174	0.175	-0.6	92	0.00
14 T	Allyl chloride	1.094	1.107	-1.2	91	0.00
15 T	Acrylonitrile	0.404	0.409	-1.2	91	0.00
16 T	Acetone	0.369	0.353	4.3	91	0.00
17 T	Carbon Disulfide	1.636	1.461	10.7	80	0.00
18 T	Methyl Acetate	0.888	0.936	-5.4	98	0.00
19 T	Methyl tert-butyl Ether	2.061	2.022	1.9	88	0.00
20 T	Methylene Chloride	0.731	0.704	3.7	92	0.00
21 T	trans-1,2-Dichloroethene	0.607	0.617	-1.6	89	0.00
22 T	Diisopropyl ether	2.224	2.283	-2.7	91	0.00
23 T	Vinyl Acetate	1.856	1.907	-2.7	88	0.00
24 P	1,1-Dichloroethane	1.247	1.229	1.4	90	0.00
25 T	2-Butanone	0.555	0.566	-2.0	89	0.00
26 T	2,2-Dichloropropane	0.585	0.730	-24.8	111	0.00
27 T	cis-1,2-Dichloroethene	0.749	0.750	-0.1	90	0.00
28 T	Bromochloromethane	0.594	0.610	-2.7	92	0.00
29 T	Tetrahydrofuran	0.373	0.364	2.4	89	0.00
30 C	Chloroform	1.239	1.232	0.6#	90	0.00
31 T	Cyclohexane	1.082	1.102	-1.8	90	0.00
32 T	1,1,1-Trichloroethane	1.000	0.986	1.4	88	0.00
33 S	1,2-Dichloroethane-d4	0.795	0.805	-1.3	97	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.334	0.344	-3.0	94	0.00
36 T	1,1-Dichloropropene	0.459	0.457	0.4	88	0.00
37 T	Ethyl Acetate	0.591	0.583	1.4	87	0.00
38 T	Carbon Tetrachloride	0.465	0.453	2.6	87	0.00
39 T	Methylcyclohexane	0.549	0.590	-7.5	91	0.00
40 TM	Benzene	1.448	1.483	-2.4	89	0.00
41 T	Methacrylonitrile	0.319	0.338	-6.0	90	0.00
42 TM	1,2-Dichloroethane	0.521	0.525	-0.8	90	0.00
43 T	Isopropyl Acetate	0.874	0.889	-1.7	86	0.00
44 TM	Trichloroethene	0.340	0.337	0.9	89	0.00
45 C	1,2-Dichloropropane	0.372	0.364	2.2#	89	0.00
46 T	Dibromomethane	0.263	0.269	-2.3	90	0.00
47 T	Bromodichloromethane	0.516	0.516	0.0	89	0.00
48 T	Methyl methacrylate	0.438	0.456	-4.1	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045153.D
 Acq On : 05 Mar 2025 19:24
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 06 01:12:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	91	0.00
50 S	Toluene-d8	1.212	1.246	-2.8	93	0.00
51 T	4-Methyl-2-Pentanone	0.587	0.614	-4.6	91	0.00
52 CM	Toluene	0.849	0.889	-4.7#	89	0.00
53 T	t-1,3-Dichloropropene	0.431	0.454	-5.3	87	0.00
54 T	cis-1,3-Dichloropropene	0.503	0.527	-4.8	89	0.00
55 T	1,1,2-Trichloroethane	0.350	0.356	-1.7	90	0.00
56 T	Ethyl methacrylate	0.532	0.558	-4.9	88	0.00
57 T	1,3-Dichloropropane	0.605	0.619	-2.3	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.297	-14.7	94	0.00
59 T	2-Hexanone	0.425	0.455	-7.1	92	0.00
60 T	Dibromochloromethane	0.366	0.373	-1.9	88	0.00
61 T	1,2-Dibromoethane	0.349	0.355	-1.7	90	0.00
62 S	4-Bromofluorobenzene	0.402	0.428	-6.5	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.320	0.324	-1.3	91	0.00
65 PM	Chlorobenzene	1.058	1.073	-1.4	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.352	0.346	1.7	89	0.00
67 C	Ethyl Benzene	1.843	1.931	-4.8#	90	0.00
68 T	m/p-Xylenes	0.675	0.716	-6.1	90	0.00
69 T	o-Xylene	0.681	0.704	-3.4	91	0.00
70 T	Styrene	1.101	1.171	-6.4	90	0.00
71 P	Bromoform	0.266	0.267	-0.4	88	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.903	4.051	-3.8	92	0.00
74 T	N-amyl acetate	1.831	1.888	-3.1	89	0.00
75 P	1,1,2,2-Tetrachloroethane	1.432	1.407	1.7	93	0.00
76 T	1,2,3-Trichloropropane	1.164	1.129	3.0	92	0.00
77 T	Bromobenzene	0.921	0.929	-0.9	93	0.00
78 T	n-propylbenzene	4.409	4.704	-6.7	92	0.00
79 T	2-Chlorotoluene	2.808	2.797	0.4	91	0.00
80 T	1,3,5-Trimethylbenzene	3.156	3.289	-4.2	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.340	0.328	3.5	91	0.00
82 T	4-Chlorotoluene	3.060	3.158	-3.2	91	0.00
83 T	tert-Butylbenzene	3.242	3.277	-1.1	91	0.00
84 T	1,2,4-Trimethylbenzene	3.175	3.311	-4.3	92	0.00
85 T	sec-Butylbenzene	3.884	4.140	-6.6	92	0.00
86 T	p-Isopropyltoluene	3.139	3.355	-6.9	92	0.00
87 T	1,3-Dichlorobenzene	1.643	1.676	-2.0	94	0.00
88 T	1,4-Dichlorobenzene	1.662	1.663	-0.1	92	0.00
89 T	n-Butylbenzene	2.703	3.020	-11.7	94	0.00
90 T	Hexachloroethane	0.568	0.550	3.2	85	0.00
91 T	1,2-Dichlorobenzene	1.648	1.677	-1.8	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.279	0.282	-1.1	91	0.00
93 T	1,2,4-Trichlorobenzene	0.938	1.001	-6.7	93	0.00
94 T	Hexachlorobutadiene	0.385	0.409	-6.2	98	0.00
95 T	Naphthalene	3.605	3.860	-7.1	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045153.D
Acq On : 05 Mar 2025 19:24
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Mar 06 01:12:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.990	1.044	-5.5	92	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6