

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
 Data File : VX045134.D
 Acq On : 05 Mar 2025 12:03
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 06 01:02:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	133109	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44724	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56875	52.871	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.740%
35) Dibromofluoromethane	5.379	113	45613	51.247	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.500%
50) Toluene-d8	8.647	98	166987	51.750	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.500%
62) 4-Bromofluorobenzene	11.079	95	52277	48.891	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.780%


Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030525\
Data File : VX045134.D
Acq On : 05 Mar 2025 12:03
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 06 01:02:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 732
Delta R.T. -0.000 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 67628
Ion Ratio Lower Upper
168 100
99 62.2 48.2 72.4



#33
1,2-Dichloroethane-d4
Concen: 52.871 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03

Tgt Ion: 65 Resp: 56875
Ion Ratio Lower Upper
65 100
67 52.4 0.0 106.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 133109
Ion Ratio Lower Upper
114 100
63 21.4 0.0 41.8
88 16.9 0.0 32.8



#35
Dibromofluoromethane
Concen: 51.247 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03

Tgt Ion:113 Resp: 45613
Ion Ratio Lower Upper
113 100
111 99.7 81.8 122.6
192 17.8 14.3 21.5









#50
Toluene-d8
Concen: 51.750 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 166987
Ion Ratio Lower Upper
98 100
100 65.7 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 48.891 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03

Tgt Ion: 95 Resp: 52277
Ion Ratio Lower Upper
95 100
174 73.8 0.0 148.2
176 70.3 0.0 141.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1470
Delta R.T. 0.006 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion:117 Resp: 119419
Ion Ratio Lower Upper
117 100
82 58.2 46.3 69.5
119 31.4 25.7 38.5



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.006 min
Lab File: VX045134.D
Acq: 05 Mar 2025 12:03

Tgt Ion:152 Resp: 44724
Ion Ratio Lower Upper
152 100
115 62.8 44.2 132.6
150 157.2 0.0 349.0

