

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030619\
 Data File : VX007864.D
 Acq On : 06 Mar 2019 17:07
 Operator : JC/SP
 Sample : K1759-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-EB-20190305

Quant Time: Mar 07 08:31:49 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	304879	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	474128	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	444457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	211898	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	166439	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	
35) Dibromofluoromethane	5.51	113	153128	50.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.98%	
50) Toluene-d8	8.71	98	572477	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	11.13	95	208780	53.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.92%	
Target Compounds						
16) Acetone	2.45	43	10946	5.880	ug/l	96
20) Methylene Chloride	2.86	84	2433	0.821	ug/l	89
25) 2-Butanone	4.70	43	3139	1.191	ug/l #	89
52) Toluene	8.79	92	4495	0.576	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030619\
Data File : VX007864.D
Acq On : 06 Mar 2019 17:07
Operator : JC/SP
Sample : K1759-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-EB-20190305

Quant Time: Mar 07 08:31:49 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

