

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030619\
 Data File : VX007869.D
 Acq On : 06 Mar 2019 19:20
 Operator : JC/SP
 Sample : K1759-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB201-GW-580-582

Quant Time: Mar 07 08:49:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	303634	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	473623	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	441028	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	212211	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	167798	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
35) Dibromofluoromethane	5.51	113	155115	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	8.71	98	571888	51.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.80%	
62) 4-Bromofluorobenzene	11.13	95	208984	53.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.14%	
Target Compounds						
16) Acetone	2.45	43	21551	11.624	ug/l	99
25) 2-Butanone	4.71	43	8520	3.246	ug/l	98
52) Toluene	8.78	92	9451	1.212	ug/l	98
64) Tetrachloroethene	9.34	164	1501	0.411	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030619\
Data File : VX007869.D
Acq On : 06 Mar 2019 19:20
Operator : JC/SP
Sample : K1759-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB201-GW-580-582

Quant Time: Mar 07 08:49:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

