

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030619\
 Data File : VX007857.D
 Acq On : 06 Mar 2019 12:20
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Mar 07 07:32:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 06:50:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	305342	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	452476	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	394075	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	182883	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	161460	50.13	ug/l	0.00
Spiked Amount			Recovery	=	100.26%	
35) Dibromofluoromethane	5.51	113	138775	47.95	ug/l	0.00
Spiked Amount			Recovery	=	95.90%	
50) Toluene-d8	8.71	98	517458	48.68	ug/l	0.00
Spiked Amount			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	11.13	95	176170	47.27	ug/l	0.00
Spiked Amount			Recovery	=	94.54%	


Target Compounds	Qvalue
------------------	--------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030619\
Data File : VX007857.D
Acq On : 06 Mar 2019 12:20
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Mar 07 07:32:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 06:50:53 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.00 min

Lab File: VX007857.D

Acq: 06 Mar 2019 12:20

Instrument :

MSVOA_X

ClientSampleId :


IBLK

Tot Ion:168 Resp: 305342

Ion Ratio Lower Upper

168 100

99 48.0 37.6 56.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

120000

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 50.126 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX007857.D


Acq: 06 Mar 2019 12:20

Tgt Ion: 65 Resp: 161460

Ion Ratio Lower Upper

65 100

67 54.4 0.0 110.0



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. 0.00 min

Lab File: VX007857.D

Acq: 06 Mar 2019 12:20

Instrument :

MSVOA_X

ClientSampleId :

IBLK


Tot Ion:114 Resp: 452476

Ion Ratio Lower Upper

114 100

63 20.2 0.0 39.8




88 14.9 0.0 31.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 47.947 ug/l

RT: 5.51 min Scan# 714

Delta R.T. 0.01 min

Lab File: VX007857.D

Acq: 06 Mar 2019 12:20


Tgt Ion:113 Resp: 138775

Ion Ratio Lower Upper

113 100

111 103.1 81.4 122.0



192 24.3 19.4 29.0




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 48.681 ug/l

RT: 8.71 min Scan# 1240

Delta R.T. 0.00 min

Lab File: VX007857.D

Acq: 06 Mar 2019 12:20

Instrument :

MSVOA_X

ClientSampleId :


IBLK

Tot Ion: 98 Resp: 517458

Ion Ratio Lower Upper

98 100

100 65.1 51.7 77.5



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

8.71

300000



200000

100000

0

Time-->

8.60 8.70 8.80



#62

4-Bromofluorobenzene

Concen: 47.272 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. 0.00 min

Lab File: VX007857.D

Acq: 06 Mar 2019 12:20


Tgt Ion: 95 Resp: 176170

Ion Ratio Lower Upper

95 100

174 91.8 0.0 182.8

176 89.0 0.0 178.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

11.13

150000


100000


50000

0

Time-->

11.10 11.20





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. 0.00 min

Lab File: VX007857.D

Acq: 06 Mar 2019 12:20

Instrument :

MSVOA_X

ClientSampled :

IBLK


Tot Ion:117 Resp: 394075

Ion Ratio Lower Upper

117 100

82 55.2 44.3 66.5




119 32.4 25.3 37.9



Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX007857.D

Acq: 06 Mar 2019 12:20


Tgt Ion:152 Resp: 182883

Ion Ratio Lower Upper

152 100

115 55.0 38.6 115.7

150 155.4 0.0 348.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

