

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
 Data File : VX021006.D
 Acq On : 04 Mar 2021 11:49
 Operator : JC/MD
 Sample : VX0306WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0306WBL01

Quant Time: Mar 04 13:35:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.635	168	219905	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.829	114	411812	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	396634	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	172403	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	148715	50.07	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.14%	
35) Dibromofluoromethane	5.464	113	129145	51.61	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.22%	
50) Toluene-d8	8.694	98	515286	52.82	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.64%	
62) 4-Bromofluorobenzene	11.118	95	185087	51.31	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
Data File : VX021006.D
Acq On : 04 Mar 2021 11:49
Operator : JC/MD
Sample : VX0306WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0306WBL01

Quant Time: Mar 04 13:35:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

