

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
 Data File : VX021010.D
 Acq On : 04 Mar 2021 13:42
 Operator : JC/MD
 Sample : M1552-06 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 435-416

Quant Time: Mar 04 14:33:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.623	168	221499	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	418323	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	389735	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.059	152	172317	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.023	65	151079	50.50	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.00%	
35) Dibromofluoromethane	5.464	113	131860	51.88	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.76%	
50) Toluene-d8	8.694	98	507836	51.24	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.48%	
62) 4-Bromofluorobenzene	11.118	95	184381	50.32	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 100.64%	
Target Compounds						
16) Acetone	2.417	43	2941	2.39	ug/l	88
20) Methylene Chloride	2.834	84	1128	0.38	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
Data File : VX021010.D
Acq On : 04 Mar 2021 13:42
Operator : JC/MD
Sample : M1552-06 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
435-416

Quant Time: Mar 04 14:33:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

