

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
 Data File : VX021015.D
 Acq On : 04 Mar 2021 15:36
 Operator : JC/MD
 Sample : M1552-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 435-416

Quant Time: Mar 05 01:22:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 24 13:50:55 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.629	168	226085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.829	114	416282	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.094	117	396529	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.065	152	168250	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.029	65	151158	49.50	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 99.00%			
35) Dibromofluoromethane	5.464	113	128897	50.96	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.92%			
50) Toluene-d8	8.694	98	518294	52.56	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.12%			
62) 4-Bromofluorobenzene	11.118	95	185163	50.78	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.56%			
Target Compounds						Qvalue
11) Tert butyl alcohol	3.011	59	3474	5.11	ug/l	96
16) Acetone	2.417	43	24511	19.48	ug/l	99
20) Methylene Chloride	2.835	84	2495	0.83	ug/l #	89
25) 2-Butanone	4.635	43	5491	2.87	ug/l	98
51) 4-Methyl-2-Pentanone	8.618	43	9456	2.54	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
Data File : VX021015.D
Acq On : 04 Mar 2021 15:36
Operator : JC/MD
Sample : M1552-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
435-416

Quant Time: Mar 05 01:22:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 24 13:50:55 2021
Response via : Initial Calibration

