

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
 Data File : VX021014.D
 Acq On : 04 Mar 2021 15:13
 Operator : JC/MD
 Sample : M1553-07 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 30-31

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M

Title : SW846 8260

Signal : TIC: VX021014.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	5.464	732	744	758	rBV2	141963	408317	31.48%	5.748%
2	5.629	760	772	787	rBV	225557	626589	48.31%	8.821%
3	6.035	828	841	855	rBV	155130	413701	31.90%	5.824%
4	6.829	965	976	991	rBV	379563	905472	69.81%	12.746%
5	8.694	1286	1293	1302	rBV	827718	1296966	100.00%	18.258%
6	10.094	1525	1531	1549	rBV	847300	1117043	86.13%	15.725%
7	11.118	1700	1705	1715	rVB	629680	829417	63.95%	11.676%
8	12.059	1859	1865	1873	rBV	765376	964607	74.37%	13.579%
9	12.141	1873	1879	1885	rVB	120957	151361	11.67%	2.131%
10	12.253	1893	1898	1927	rBV	164961	322740	24.88%	4.543%
11	13.553	2113	2119	2125	rBV	13347	21059	1.62%	0.296%
12	13.871	2168	2173	2180	rBV	20893	31738	2.45%	0.447%
13	16.777	2661	2667	2669	rBV5	8461	14709	1.13%	0.207%




Sum of corrected areas: 7103719

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
Data File : VX021014.D
Acq On : 04 Mar 2021 15:13
Operator : JC/MD
Sample : M1553-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30-31

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
Data File : VX021014.D
Acq On : 04 Mar 2021 15:13
Operator : JC/MD
Sample : M1553-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30-31










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 1 Eucalyptol Concentration Rank 2

R.T.	EstConc	Area	Relative to ISTD	R.T.
12.141	7.85 ug/l	151361	1,4-Dichlorobenzene-d4	12.059

Hit#	of	5	Tentative ID	MW	MolForm	CAS#	Qual
1			Eucalyptol	154	C10H18O	000470-82-6	99
2			Trifluoroacetyl-.alpha.-terpineol	250	C12H17F3O2	1000058-17-6	74
3			2-Isopropyl-5-methylhex-2-enal	154	C10H18O	035158-25-9	22
4			2-Acetylcyclopentanone	126	C7H10O2	001670-46-8	16
5			3-Cyclohexen-1-ol, 4-methyl-1-(1...	154	C10H18O	020126-76-5	16



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
Data File : VX021014.D
Acq On : 04 Mar 2021 15:13
Operator : JC/MD
Sample : M1553-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30-31










Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

Peak Number 2 1-Hexanol, 2-ethyl- Concentration Rank 1

R.T.	EstConc	Area	Relative to ISTD	R.T.
12.253	16.73 ug/l	322740	1,4-Dichlorobenzene-d4	12.059

Hit#	of 5	Tentative ID	MW	MolForm	CAS#	Qual
1		1-Hexanol, 2-ethyl-	130	C8H18O	000104-76-7	78
2		1-Pentanol, 2-ethyl-4-methyl-	130	C8H18O	000106-67-2	56
3		Heptane, 1,1'-oxybis-	214	C14H30O	000629-64-1	53
4		Decane, 2,6,7-trimethyl-	184	C13H28	062108-25-2	53
5		2-Propyl-1-pentanol	130	C8H18O	058175-57-8	47



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030621\
Data File : VX021014.D
Acq On : 04 Mar 2021 15:13
Operator : JC/MD
Sample : M1553-07 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
30-31

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022421W.M
Quant Title : SW846 8260

TIC Library : C:\DATABASE\NIST11.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc
Eucalyptol	12.141	7.8	ug/l	151361	4	12.059	964607	50.0
1-Hexanol, 2-et...	12.253	16.7	ug/l	322740	4	12.059	964607	50.0