

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030624\
 Data File : VX040596.D
 Acq On : 06 Mar 2024 10:31
 Operator : JC/MD
 Sample : VX0306WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0306WBL01

Quant Time: Mar 07 04:33:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Mar 05 13:21:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.903	128	15028	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	85442	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	79404	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.958	65	45353	30.217	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	100.733%
60) 4-Bromofluorobenzene	11.079	95	38767	28.463	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	94.867%
63) Toluene-d8	8.646	98	103485	29.491	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030624\
Data File : VX040596.D
Acq On : 06 Mar 2024 10:31
Operator : JC/MD
Sample : VX0306WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0306WBL01

Quant Time: Mar 07 04:33:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X030524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Mar 05 13:21:29 2024
Response via : Initial Calibration

