

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045157.D
 Acq On : 06 Mar 2025 10:46
 Operator : JC/MD
 Sample : VX0306WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0306WBL01

Quant Time: Mar 07 00:37:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	71177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	139793	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	129041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53719	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	61044	53.917	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.840%
35) Dibromofluoromethane	5.379	113	48946	52.362	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.720%
50) Toluene-d8	8.647	98	177288	52.315	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.640%
62) 4-Bromofluorobenzene	11.079	95	58160	51.792	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045157.D
Acq On : 06 Mar 2025 10:46
Operator : JC/MD
Sample : VX0306WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0306WBL01

Quant Time: Mar 07 00:37:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

