

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045159.D
 Acq On : 06 Mar 2025 11:36
 Operator : JC/MD
 Sample : Q1501-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16B-87.5-030425

Quant Time: Mar 07 01:36:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	72795	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	143393	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59112	51.051	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.100%
35) Dibromofluoromethane	5.385	113	47786	49.838	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.680%
50) Toluene-d8	8.646	98	175482	50.482	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.960%
62) 4-Bromofluorobenzene	11.079	95	56245	48.829	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.660%
Target Compounds						
16) Acetone	2.385	43	6724	12.516	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045159.D
Acq On : 06 Mar 2025 11:36
Operator : JC/MD
Sample : Q1501-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16B-87.5-030425

Quant Time: Mar 07 01:36:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

