

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045160.D
 Acq On : 06 Mar 2025 12:00
 Operator : JC/MD
 Sample : Q1501-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-16B-87.5-030425-FD

Quant Time: Mar 07 00:39:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	75969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138477	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59093	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62783	51.956	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.920%	
35) Dibromofluoromethane	5.385	113	51551	51.314	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.620%	
50) Toluene-d8	8.647	98	190091	52.193	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.380%	
62) 4-Bromofluorobenzene	11.079	95	64236	53.225	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 106.440%	
Target Compounds						
16) Acetone	2.380	43	7022	12.525	ug/l	Qvalue # 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045160.D
Acq On : 06 Mar 2025 12:00
Operator : JC/MD
Sample : Q1501-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-16B-87.5-030425-FD

Quant Time: Mar 07 00:39:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

