

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045165.D
 Acq On : 06 Mar 2025 13:56
 Operator : JC/MD
 Sample : Q1484-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-OTR-COMP-01

Quant Time: Mar 07 00:42:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69378	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	140193	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	124536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	46314	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59870	54.252	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.500%
35) Dibromofluoromethane	5.379	113	47831	51.023	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.040%
50) Toluene-d8	8.647	98	175688	51.695	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.400%
62) 4-Bromofluorobenzene	11.079	95	55872	49.612	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.220%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	39316	76.786	ug/l	96
18) Methyl Acetate	2.703	43	3784	3.072	ug/l	94
20) Methylene Chloride	2.788	84	2611	2.574	ug/l #	94
43) Isopropyl Acetate	6.342	43	47179	19.260	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045165.D
Acq On : 06 Mar 2025 13:56
Operator : JC/MD
Sample : Q1484-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-OTR-COMP-01

Quant Time: Mar 07 00:42:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

