

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045166.D
 Acq On : 06 Mar 2025 14:19
 Operator : JC/MD
 Sample : Q1494-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PURGE-WATER

Quant Time: Mar 07 00:43:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	75312	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.756	114	147304	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	136609	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	58539	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	63093	52.668	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.340%	
35) Dibromofluoromethane	5.385	113	51608	52.395	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.780%	
50) Toluene-d8	8.646	98	182955	51.235	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.460%	
62) 4-Bromofluorobenzene	11.079	95	63705	53.837	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 107.680%	
Target Compounds						Qvalue
16) Acetone	2.379	43	5757	10.358	ug/l	98
18) Methyl Acetate	2.708	43	1605	1.200	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045166.D
Acq On : 06 Mar 2025 14:19
Operator : JC/MD
Sample : Q1494-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PURGE-WATER

Quant Time: Mar 07 00:43:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

