

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045169.D
 Acq On : 06 Mar 2025 15:29
 Operator : JC/MD
 Sample : Q1481-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-K1311

Quant Time: Mar 07 00:45:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 03/10/2025
 Supervised By : Mahesh Dadoda 03/10/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	72948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147727	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	136246	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63244	54.504	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.000%			
35) Dibromofluoromethane	5.379	113	51625	52.262	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 104.520%			
50) Toluene-d8	8.647	98	184918	51.636	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 103.280%			
62) 4-Bromofluorobenzene	11.079	95	62816	52.934	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 105.860%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	39933	74.175	ug/l	97
18) Methyl Acetate	2.703	43	1698	1.311	ug/l	92
20) Methylene Chloride	2.788	84	2512	2.356	ug/l	96
25) 2-Butanone	4.580	43	4207m	5.191	ug/l	
43) Isopropyl Acetate	6.336	43	47894	18.555	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045169.D
Acq On : 06 Mar 2025 15:29
Operator : JC/MD
Sample : Q1481-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-K1311

Quant Time: Mar 07 00:45:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 03/10/2025
Supervised By : Mahesh Dadoda 03/10/2025

