

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045170.D
 Acq On : 06 Mar 2025 15:52
 Operator : JC/MD
 Sample : Q1488-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ENV-101-SB01

Quant Time: Mar 07 00:45:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	78735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	160534	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59867	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68170	54.432	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.860%	
35) Dibromofluoromethane	5.379	113	55628	51.822	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.640%	
50) Toluene-d8	8.647	98	201436	51.761	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.520%	
62) 4-Bromofluorobenzene	11.079	95	67095	52.029	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.060%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	41771	71.886	ug/l	100
18) Methyl Acetate	2.709	43	2222	1.590	ug/l	97
20) Methylene Chloride	2.788	84	2646	2.299	ug/l	89
25) 2-Butanone	4.587	43	4567	5.221	ug/l	97
43) Isopropyl Acetate	6.342	43	50668	18.063	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045170.D
Acq On : 06 Mar 2025 15:52
Operator : JC/MD
Sample : Q1488-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ENV-101-SB01

Quant Time: Mar 07 00:45:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

