

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045171.D
 Acq On : 06 Mar 2025 16:16
 Operator : JC/MD
 Sample : Q1488-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ENV-101-SB02

Quant Time: Mar 07 00:46:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	76620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	153454	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	144581	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66538	54.595	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.200%	
35) Dibromofluoromethane	5.385	113	53501	52.140	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.280%	
50) Toluene-d8	8.647	98	197961	53.215	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.440%	
62) 4-Bromofluorobenzene	11.079	95	64881	52.633	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.260%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	44531	78.751	ug/l	97
18) Methyl Acetate	2.703	43	2504	1.841	ug/l	99
20) Methylene Chloride	2.788	84	2398	2.141	ug/l	90
25) 2-Butanone	4.593	43	5173	6.078	ug/l #	88
43) Isopropyl Acetate	6.342	43	49863	18.597	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045171.D
Acq On : 06 Mar 2025 16:16
Operator : JC/MD
Sample : Q1488-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ENV-101-SB02

Quant Time: Mar 07 00:46:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

