

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030625\  
 Data File : VX045172.D  
 Acq On : 06 Mar 2025 16:39  
 Operator : JC/MD  
 Sample : Q1488-06  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 ENV-102-SB01

Quant Time: Mar 07 00:46:58 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022825W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Feb 28 06:45:16 2025  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|------------|----------|
| -----                       |           |       |          |          |            |          |
| Internal Standards          |           |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.550     | 168   | 84608    | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.757     | 114   | 167928   | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.049    | 117   | 153445   | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018    | 152   | 63036    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |           |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.952     | 65    | 72361    | 53.767   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 74 - 125 | Recovery | = 107.540% |          |
| 35) Dibromofluoromethane    | 5.379     | 113   | 57808    | 51.481   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 75 - 124 | Recovery | = 102.960% |          |
| 50) Toluene-d8              | 8.647     | 98    | 208160   | 51.134   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 86 - 113 | Recovery | = 102.260% |          |
| 62) 4-Bromofluorobenzene    | 11.079    | 95    | 70449    | 52.225   | ug/l       | 0.00     |
| Spiked Amount               | 50.000    | Range | 77 - 121 | Recovery | = 104.440% |          |
| Target Compounds            |           |       |          |          |            |          |
|                             |           |       |          |          | Qvalue     |          |
| 16) Acetone                 | 2.380     | 43    | 43675    | 69.945   | ug/l       | 99       |
| 18) Methyl Acetate          | 2.709     | 43    | 3472     | 2.311    | ug/l       | 99       |
| 20) Methylene Chloride      | 2.782     | 84    | 2563     | 2.072    | ug/l #     | 90       |
| 25) 2-Butanone              | 4.581     | 43    | 5867     | 6.242    | ug/l       | 93       |
| 43) Isopropyl Acetate       | 6.342     | 43    | 49551    | 16.887   | ug/l       | 98       |
| -----                       |           |       |          |          |            |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX030625\  
Data File : VX045172.D  
Acq On : 06 Mar 2025 16:39  
Operator : JC/MD  
Sample : Q1488-06  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
ENV-102-SB01

Quant Time: Mar 07 00:46:58 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X022825W.M  
Quant Title : SW846 8260  
QLast Update : Fri Feb 28 06:45:16 2025  
Response via : Initial Calibration

