

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045174.D
 Acq On : 06 Mar 2025 17:25
 Operator : JC/MD
 Sample : Q1488-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ENV-104-SB01

Quant Time: Mar 07 00:47:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	75491	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152473	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	55464	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63445	52.836	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.680%
35) Dibromofluoromethane	5.379	113	51588	50.599	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.200%
50) Toluene-d8	8.647	98	190820	51.626	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.260%
62) 4-Bromofluorobenzene	11.079	95	63521	51.862	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.720%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	41961	75.316	ug/l	95
18) Methyl Acetate	2.703	43	3776	2.817	ug/l	98
20) Methylene Chloride	2.788	84	2526	2.289	ug/l	89
25) 2-Butanone	4.574	43	4523	5.393	ug/l	97
43) Isopropyl Acetate	6.342	43	42994	16.138	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045174.D
Acq On : 06 Mar 2025 17:25
Operator : JC/MD
Sample : Q1488-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ENV-104-SB01

Quant Time: Mar 07 00:47:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

