

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045177.D
 Acq On : 06 Mar 2025 18:35
 Operator : JC/MD
 Sample : Q1494-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ASPHALT-SOIL

Quant Time: Mar 07 00:49:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	76042	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150194	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	54345	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64892	53.649	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.300%
35) Dibromofluoromethane	5.379	113	52345	52.120	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.240%
50) Toluene-d8	8.647	98	190206	52.241	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.480%
62) 4-Bromofluorobenzene	11.079	95	62120	51.487	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.980%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	19854	35.378	ug/l	100
18) Methyl Acetate	2.703	43	3136	2.323	ug/l	99
20) Methylene Chloride	2.788	84	3460	3.112	ug/l #	93
43) Isopropyl Acetate	6.342	43	35177	13.404	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045177.D
Acq On : 06 Mar 2025 18:35
Operator : JC/MD
Sample : Q1494-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ASPHALT-SOIL

Quant Time: Mar 07 00:49:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

