

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
 Data File : VX045180.D
 Acq On : 06 Mar 2025 19:45
 Operator : JC/MD
 Sample : Q1478-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-SO-DRUM-582-022825

Quant Time: Mar 07 00:51:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 28 06:45:16 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	74182	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	149089	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	138397	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	57244	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62937	53.338	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.680%	
35) Dibromofluoromethane	5.385	113	50362	50.518	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.040%	
50) Toluene-d8	8.647	98	187727	51.942	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.880%	
62) 4-Bromofluorobenzene	11.079	95	62997	52.601	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 105.200%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	41661	76.097	ug/l	97
18) Methyl Acetate	2.703	43	2526	1.918	ug/l	93
20) Methylene Chloride	2.782	84	4062	3.746	ug/l	93
25) 2-Butanone	4.586	43	4292	5.208	ug/l	98
43) Isopropyl Acetate	6.342	43	46947	18.022	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX030625\
Data File : VX045180.D
Acq On : 06 Mar 2025 19:45
Operator : JC/MD
Sample : Q1478-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-SO-DRUM-582-022825

Quant Time: Mar 07 00:51:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X022825W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 28 06:45:16 2025
Response via : Initial Calibration

