

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007880.D
 Acq On : 07 Mar 2019 13:38
 Operator : JC/SP
 Sample : K1781-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190305

Quant Time: Mar 08 05:50:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	291580	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	453796	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	427994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	207827	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161621	52.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.08%	
35) Dibromofluoromethane	5.51	113	147873	50.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.88%	
50) Toluene-d8	8.71	98	550446	51.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.26%	
62) 4-Bromofluorobenzene	11.13	95	200829	53.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.46%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007880.D
Acq On : 07 Mar 2019 13:38
Operator : JC/SP
Sample : K1781-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190305

Quant Time: Mar 08 05:50:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

