

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007888.D
 Acq On : 07 Mar 2019 17:11
 Operator : JC/SP
 Sample : K1783-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-3

Quant Time: Mar 08 06:39:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	288242	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	451232	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	421176	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200320	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	158600	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
35) Dibromofluoromethane	5.50	113	149350	51.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.48%	
50) Toluene-d8	8.71	98	544552	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
62) 4-Bromofluorobenzene	11.13	95	196786	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
Target Compounds						
16) Acetone	2.45	43	5422	3.081	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007888.D
Acq On : 07 Mar 2019 17:11
Operator : JC/SP
Sample : K1783-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-3

Quant Time: Mar 08 06:39:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

