

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007892.D
 Acq On : 07 Mar 2019 18:58
 Operator : JC/SP
 Sample : K1793-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D3-20190306

Quant Time: Mar 08 06:53:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	285418	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	449131	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	418251	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200803	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	159329	52.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.84%	
35) Dibromofluoromethane	5.50	113	149468	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
50) Toluene-d8	8.71	98	543666	51.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.06%	
62) 4-Bromofluorobenzene	11.13	95	197120	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.38	96	524	0.216	ug/l	93
16) Acetone	2.45	43	4818	2.765	ug/l	98
44) Trichloroethene	7.21	130	12900	3.866	ug/l	99
64) Tetrachloroethene	9.34	164	7377	2.128	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007892.D
Acq On : 07 Mar 2019 18:58
Operator : JC/SP
Sample : K1793-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D3-20190306

Quant Time: Mar 08 06:53:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

