

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007893.D
 Acq On : 07 Mar 2019 19:24
 Operator : JC/SP
 Sample : K1793-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-20190306

Quant Time: Mar 08 06:57:16 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	280402	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	439704	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	414686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	196108	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	157427	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
35) Dibromofluoromethane	5.50	113	145697	51.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.60%	
50) Toluene-d8	8.71	98	533696	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.13	95	195130	53.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.76%	
Target Compounds						
44) Trichloroethene	7.21	130	136514	41.793	ug/l	99
49) 1,4-Dioxane	7.76	88	913	12.089	ug/l	93
64) Tetrachloroethene	9.34	164	4904	1.427	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007893.D
Acq On : 07 Mar 2019 19:24
Operator : JC/SP
Sample : K1793-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-20190306

Quant Time: Mar 08 06:57:16 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

