

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007901.D
 Acq On : 07 Mar 2019 23:50
 Operator : JC/SP
 Sample : PB117680TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB117680TB

Quant Time: Mar 08 07:20:23 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	281783	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	441868	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	424822	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200323	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	156259	52.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.14%	
35) Dibromofluoromethane	5.50	113	144964	51.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.58%	
50) Toluene-d8	8.71	98	546290	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
62) 4-Bromofluorobenzene	11.13	95	199985	54.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.90%	
Target Compounds						
16) Acetone	2.45	43	12624	7.337	ug/l	98
20) Methylene Chloride	2.86	84	7328	2.676	ug/l	97
25) 2-Butanone	4.70	43	2781	1.142	ug/l	94
43) Isopropyl Acetate	6.46	43	17319	2.465	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	7580	0.680	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007901.D
Acq On : 07 Mar 2019 23:50
Operator : JC/SP
Sample : PB117680TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB117680TB

Quant Time: Mar 08 07:20:23 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

