

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX030719\
 Data File : VX007905.D
 Acq On : 08 Mar 2019 01:37
 Operator : JC/SP
 Sample : K1753-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z-2-6

Quant Time: Mar 08 08:00:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 07 07:33:26 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	288190	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	450621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	431007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	207525	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	159510	52.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.94%	
35) Dibromofluoromethane	5.50	113	147591	51.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.40%	
50) Toluene-d8	8.71	98	556053	52.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.06%	
62) 4-Bromofluorobenzene	11.13	95	201681	54.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.68%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.06	59	2414	3.551	ug/l	99
16) Acetone	2.45	43	9706	5.516	ug/l	99
18) Methyl Acetate	2.78	43	7843	2.098	ug/l	97
20) Methylene Chloride	2.86	84	17896	6.390	ug/l	95
25) 2-Butanone	4.70	43	6380	2.561	ug/l	99
43) Isopropyl Acetate	6.46	43	17079	2.383	ug/l	99
95) Naphthalene	13.83	128	29628	2.164	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX030719\
Data File : VX007905.D
Acq On : 08 Mar 2019 01:37
Operator : JC/SP
Sample : K1753-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z-2-6

Quant Time: Mar 08 08:00:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X030619W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 07 07:33:26 2019
Response via : Initial Calibration

